

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554



QUALITY CONTROL CERTIFICATE

KIT NAME Epstein-Barr VCA IgG - CHORUS

Code 81055

Lot No. 019

2025-06

2024-03

REAGENTS

Label	Description	Lot	Exp.	NO
D2341	Epstein-Barr VCA IgG	019	2025-06	6X6
B2342	CALIBRATORE Epstein-Barr VCA IgG	019	2025-06	1
C2341	CONTROL SERUM Epstein-Barr VCA IgG	019	2025-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	310	≥ 100
Control Serum (INDEX)	3,2	2,5-4,7
Negative Control (INDEX)	0,1	$< 0,7$


CALIBRATOR

0 0 5 0 1 9 7 0 0 0 0 0 2 8 0 0 0 0 0

Label checked with barcode reader

YES

Operator Signature *MS*



CONTROL SERUM

0 0 5 0 1 9 7 2 5 4 7 0 0 0 0 0 0 0 1

Label checked with barcode reader

YES

Operator Signature *MS*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *MS*

DATE 28.06.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.