



QUALITY CONTROL CERTIFICATE

KIT NAME: SANITIZING SOLUTION

|         |       |        |         |
|---------|-------|--------|---------|
| Code    | 83604 |        |         |
| Lot No. | 066   | Expiry | 2024-07 |

REAGENTS

| Label | Description         | Lot. | Exp     | No |
|-------|---------------------|------|---------|----|
| B2563 | SANITIZING SOLUTION | 066  | 2024-07 | 1  |

CONTROLS

|         | Found results | Expected results |
|---------|---------------|------------------|
| Density | 1.061         | 1.060-1.100      |
| pH      | 13,4          | 12-14            |

The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi *Simone Baiocchi*  
 APPROVED \_\_\_\_\_ RESPONSIBLE HEAD, Quality Control Reagents... *[Signature]*  
 DATE: *16.08.2021*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.

na