

DIESSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
 Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000-T3 COMMON SILVER BIRCH

Code	88000-T3		
Lot No.	817	2023-03	2022-03

REAGENTS

Label	Description	Lot	Exp	No
D4971	88000-T3 COMMON SILVER BIRCH	817	2023-03	2X6
B4970	CALIBRATORE 88000-T3 COMMON SILVER BIRCH	817	2023-03	1
C4970	CONTROL SERUM 88000-T3 COMMON SILVER BIRCH	817	2023-03	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	280	≥ 100
Control Serum (KU/L)	12,8	7,9-16,2

CALIBRATOR

6	0	5	8	1	7	6	5	1	7	6	4	2	2	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SL*



CONTROL SERUM

6	0	5	8	1	7	6	3	1	5	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SL*



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo *Antonella Leo*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents... *SL*

DATE *16.03.2022*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.