



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS LISTERIA MONOCYTOGENES CONTROL SERUM

Code	85875
Lot. No 079	 2024-02  2023-02

REAGENTS

CODE	LABEL	DESCRIPTION	Lot No.	EXP	NO
PF 90875	B2980	CHORUS LISTERIA MONOCYTOGENES	079	2024-02	1

CONTROLS

TESTED WITH:	FOUND RESULTS	EXPECTED RESULTS
CHORUS LISTERIA MONOCYTOGENES LOT 902	1/16	1/16 ± 1 Dil.

The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED BY:

(RESPONSIBLE HEAD, Quality Control Reagents)

DATE 24/03/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.