

QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA PNEUMONIAE IgM - CHORUS

Code	81035				
Lot No.	917		2024-06		2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2261	MYCOPLASMA PNEUMONIAE IgM	917	2024-06	6x6
B2261	CALIBRATORE MYCOPLASMA PNEUMONIAE IgM	917	2024-06	1
C2261	CONTROL SERUM MYCOPLASMA PNEUMONIAE IgM	917	2024-06	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	329	≥ 100
Control Serum (INDEX)	1,8	1,7-3,2
Negative Control (INDEX)	0,5	≤ 0,7

CALIBRATOR

0 6 7 9 1 7 7 0 0 0 0 0 4 1 0 0 0 0

Label checked with barcode reader YES Operator Signature *FR*



CONTROL SERUM

0 6 7 9 1 7 7 1 7 3 2 0 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature *FR*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *FR*

DATE *17.01.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.