

DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME MUMPS IgG

Code	81075				
Lot No.	934		2024-08		2023-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2451	MUMPS IgG	934	2024-08	6X6
B2452	CALIBRATORE MUMPS IgG	934	2024-08	1
C2451	CONTROL SERUM MUMPS IgG	934	2024-08	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	464	≥ 100
Control Serum (Index)	2,4	1,8-3,3
Negative Control (Index)	0,0	< 0,7

CALIBRATOR

0	1	0	9	3	4	7	0	0	0	0	0	4	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



CONTROL SERUM

0	1	0	9	3	4	7	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 23/05/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent reagents should be handled with maximum attention.