



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME MUMPS IgG

Code	81075				
Lot No.	935		2024-08		2023-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2451	MUMPS IgG	935	2024-08	6X6
B2452	CALIBRATORE MUMPS IgG	935	2024-08	1
C2451	CONTROL SERUM MUMPS IgG	935	2024-08	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	460	≥ 100
Control Serum (Index)	2,3	1,8-3,3
Negative Control (Index)	0,0	< 0,7

CALIBRATOR

0 1 0 9 3 5 7 0 0 0 0 0 4 0 0 0 0 0

Label checked with barcode reader YES Operator Signature *FB*



CONTROL SERUM

,0 1 0 9 3 5 7 1 8 3 3 0 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature *FB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *[Signature]*

DATE 09/06/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent reagents should be handled with maximum attention.