

QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgM -CHORUS

Code	81021		
Lot No.	928	2024-04	2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2221	Herpes simplex 1+2 IgM	928	2024-04	6x6
B2221	CALIBRATORE Herpes simplex 1+2 IgM	928	2024-04	1
C2221	CONTROL SERUM Herpes simplex 1+2 IgM	928	2024-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	490	≥ 100
Control Serum (INDEX)	2,5	1,8-3,3
Negative Control (INDEX)	0,0	< 0,7

CALIBRATOR

1	5	6	9	2	8	6	0	0	0	0	0	4	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

1669286000040000
Herpes simplex 1+2 IgM
CALIBRATOR 0.175 mL IVD
LOT 928
2024-04 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

1	5	6	9	2	8	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

166928618330000001
Herpes simplex 1+2 IgM
CONTROL+ 0.425 mL IVD
LOT 928
2024-04 2°C-8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 13.06.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.