



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME Jo-1 - CHORUS

Code	86030		
Lot No.	944	2024-09	2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3261	Jo-1	944	2024-09	6X6
B3261	CALIBRATORE Jo-1	944	2024-09	1
C3261	CONTROL SERUM Jo-1	944	2024-09	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	430	≥ 100
Control Serum (AU/mL)	31,5	23,8-57,0

CALIBRATOR

1	9	7	9	4	4	7	1	2	2	3	2	4	2	4	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *MB*



CONTROL SERUM

1	9	7	9	4	4	7	1	7	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *MB*

DATE *20.06.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.