

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHLAMYDOPHILA PNEUMONIAE IgM - CHORUS

Code	81251
Lot No.	944
	2024-06
	2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D8600	Chlamydomphila pneumoniae IgM	944	2024-06	6X6
B8600	CALIBRATORE Chlamydomphila p. IgM	944	2024-06	1
C8600	CONTROL SERUM Chlamydomphila p. IgM	944	2024-06	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	520	≥ 100
Control Serum (Index)	2,9	2,1-3,9
Negative Control (Index)	0,1	$\leq 0,7$

CALIBRATOR

2	3	0	9	4	4	6	0	0	0	0	0	3	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *FB*



CONTROL SERUM

2	3	0	9	4	4	6	2	1	3	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE 05/07/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.