



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME **CARDIOLIPIN-M - CHORUS**

Code	86048		
Lot No.	946	2024-06	2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3331	CARDIOLIPIN-M	946	2024-06	6X6
B3331	CALIBRATORE CARDIOLIPIN-M	946	2024-06	1
C3331	CONTROL SERUM CARDIOLIPIN-M	946	2024-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	550	≥ 100
Control Serum (MPLU/mL)	46,9	28,3-77,1

CALIBRATOR

3	3	6	9	4	6	6	4	1	9	3	1	3	8	3	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*

336946641931393800
Cardiolipin-M
CALIBRATOR 0.176 mL IVD
LOT 946
2024-06 2°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

3	3	6	9	4	6	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*

33694661834000001
Cardiolipin-M
CONTROL+ 0.428 mL IVD
LOT 946
2024-06 2°C
DIESSSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE *12/07/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.