



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada del Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgM -CHORUS

Code	81021		
Lot No.	929	2024-04	2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2221	Herpes simplex 1+2 IgM	929	2024-04	6x6
B2221	CALIBRATORE Herpes simplex 1+2 IgM	929	2024-04	1
C2221	CONTROL SERUM Herpes simplex 1+2 IgM	929	2024-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	500	≥ 100
Control Serum (INDEX)	2,4	1,9-3,6
Negative Control (INDEX)	0,0	<0,7

CALIBRATOR

1	5	6	9	2	9	6	0	0	0	0	0	0	3	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *HB*



CONTROL SERUM

1	5	6	9	2	9	6	1	9	3	6	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *HB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *[Signature]*

DATE 13/07/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent reagents should be handled with maximum attention.