



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-G - CHORUS

Code	86032		
Lot No.	954	2024-07	2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D3281	dsDNA-G	954	2024-07	6X6
B3281	CALIBRATORE dsDNA-G	954	2024-07	1
C3281	CONTROL SERUM dsDNA-G	954	2024-07	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	589	≥ 100
Control Serum (IU/mL)	48,4	33,3-63,6

CALIBRATOR

2	5	1	9	5	4	6	6	2	5	3	6	4	8	4	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

261954662536484000
dsDNA-G
CALIBRATOR 0.175 mL IVD
LOT 954
2024-07 2°C-8°C B3281
DIESSE Monteriggioni Italy CE

CONTROL SERUM

2	5	1	9	5	4	6	1	5	2	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

261954616280000001
dsDNA-G
CONTROL+ 0.425 mL IVD
LOT 954
2024-07 2°C-8°C C3281
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 18-07-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.