

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-G - CHORUS

Code	86032		
Lot No.	955	2024-07	2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D3281	dsDNA-G	955	2024-07	6X6
B3281	CALIBRATORE dsDNA-G	955	2024-07	1
C3281	CONTROL SERUM dsDNA-G	955	2024-07	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	330	≥ 100
Control Serum (IU/mL)	60,5	37,5-86,4

CALIBRATOR

2	5	1	9	5	5	6	7	2	6	3	7	3	9	4	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *YB*

261956672637394100  
dsDNA-G  
CALIBRATOR 0.176 mL IVD  
LOT 955  
2024-07 2°C 8°C  
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

2	5	1	9	5	5	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *YB*

2619566818340000001  
dsDNA-G  
CONTROL+ 0.426 mL IVD  
LOT 955  
2024-07 2°C 8°C  
DIESSSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *22/07/2023*

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.