



QUALITY CONTROL CERTIFICATE

KIT NAME VARICELLA IgG - CHORUS

Code	81078		
Lot No.	946	2024-06	2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D2471	VARICELLA IgG	946	2024-06	6x6
B2472	CALIBRATORE VARICELLA IgG	946	2024-06	1
C2471	CONTROL SERUM VARICELLA IgG	946	2024-06	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	462	≥ 100
Control Serum (Index)	2,8	2,4-4,5
Negative Control (Index)	0,1	< 0,7

CALIBRATOR

0	1	4	9	4	6	6	0	0	0	0	0	2	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SB*



CONTROL SERUM

0	1	4	9	4	6	6	2	4	4	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *SB*

DATE *26.07.23*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.