



QUALITY CONTROL CERTIFICATE

KIT NAME HSV 1 SCREEN - CHORUS

Code	81018			
Lot No.	952	2024-10	2023-07	

REAGENTS

Label	Description	Lot.	Exp.	NO
D2191	HSV 1 SCREEN	952	2024-10	6x6
B2191	CALIBRATORE HSV 1 SCREEN	952	2024-10	1
C2191	CONTROL SERUM HSV 1 SCREEN	952	2024-10	1


CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	480	≥ 100
CONTROL SERUM (Index)	2,4	1,3-8,0
NEGATIVE CONTROL (Index)	0,6	< 0,7

CALIBRATOR

0	3	3	9	5	2	7	0	0	0	0	0	6	8	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature MB



CONTROL SERUM

0	3	3	9	5	2	7	1	3	8	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents MB

DATE 01/08/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.