

QUALITY CONTROL CERTIFICATE

KIT NAME TICK-BORNE ENCEPHALITIS VIRUS IgM - CHORUS

Code	81277		
Lot No.	952	2024-07	2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D0160	TICK-BORNE ENCEPHALITIS VIRUS IgM	952	2024-07	6X6
B0160	CALIBRATORE TICK-BORNE ENCEPHALITIS VIRUS IgM	952	2024-12	1
C0160	CONTROL SERUM TICK-BORNE ENCEPHALITIS VIRUS IgM	952	2024-12	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	350	≥ 100
Control (Index)	2,4	1,7-3,1

CALIBRATOR

9	3	7	9	5	2	6	0	0	0	0	0	4	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SB*

93795260000420000
Tick-borne encephalitis virus IgM
CALIBRATOR 0.480 mL IVD
LOT 952
2024-12 2°C -8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

9	3	7	9	5	2	6	1	7	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SB*

93795261731000001
Tick-borne encephalitis virus IgM
CONTROL+ 0.900 mL IVD
LOT 952
2024-12 2°C -8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE *02.08.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.