

**QUALITY CONTROL CERTIFICATE**

**KIT NAME ASCA-G - CHORUS**

Code	86064		
Lot No.	943	2024-06	2023-06

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D3491	ASCA-G	943	2024-06	6X6
B3491	CALIBRATORE ASCA-G	943	2024-06	1
C3491	CONTROL SERUM ASCA-G	943	2024-06	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	349	> 100
Control Serum (AU/mL)	67,0	50,0-136,8

**CALIBRATOR**

5	0	8	9	4	3	6	6	2	1	3	1	3	4	6	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

608943662131346900  
ASCA-G  
CALIBRATOR 0.175 mL [IVD]  
LOT 943  
2024-06 2°C-8°C  
DIESSA Monteriggioni Italy CE

**CONTROL SERUM**

5	0	8	9	4	3	6	2	1	3	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

608943621380000001  
ASCA-G  
CONTROL 0.425 mL [IVD]  
LOT 943  
2024-06 2°C-8°C  
DIESSA Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 08.08.2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.