

QUALITY CONTROL CERTIFICATE

KIT NAME BORDETELLA PERTUSSIS TOXIN IgG - CHORUS

Code	81272		
Lot No.	962	2024-08	2023-08

REAGENTS

Label	Description	Lot.	Exp	No
D2100	BORDETELLA PERTUSSIS TOXIN IgG	962	2024-08	6x6
B2100	CALIBRATORE BORDETELLA P.TOXIN IgG	962	2024-09	1
C2100	CONTROL SERUM BORDETELLA P.TOXIN IgG	962	2024-09	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	440	≥100
CONTROL SERUM (IU/ml)	148,1	100,0-250,0

CALIBRATOR

9	6	4	9	6	2	6	1	2	0	3	4	3	6	4	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

964962612034364400
Bordetella pertussis
Toxin IgG
CALIBRATOR 0.450 mL IVD
LOT 962
2024-09 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

9	6	4	9	6	2	6	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

964962620360000001
Bordetella pertussis
Toxin IgG
CONTROL+ 0.800 mL IVD
LOT 962
2024-09 2°C-8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 09/08/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.