



QUALITY CONTROL CERTIFICATE

KIT NAME West Nile IgM -CHORUS

| | | | |
|---------|-------|---------|---------|
| Code | 81176 | | |
| Lot No. | 952 | 2024-07 | 2023-07 |

REAGENTS

| Label | Description | Lot. | Exp | No |
|-------|-----------------------|------|---------|-----|
| D4090 | West Nile IgM | 952 | 2024-07 | 6X6 |
| B4090 | Calibratore West Nile | 952 | 2024-07 | 1 |
| C4090 | Control West Nile | 952 | 2024-07 | 1 |

CONTROLS

| Found results | Expected results |
|---------------|------------------|
|---------------|------------------|

| | | |
|------------------|-----|---------|
| Calibrator (mOD) | 410 | ≥ 100 |
| Control (IndexI) | 3,3 | 2,2-4,0 |

CALIBRATOR

| | | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 9 | 8 | 9 | 9 | 5 | 2 | 6 | 0 | 0 | 0 | 0 | 0 | 0 | 3 | 3 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature *MB*

98995260000330000
West Nile Virus IgM
CALIBRATOR 0.175 mL IVD
LOT 952
2024-07
DIESSE Monteriggioni Italy
CE
B4090

CONTROL SERUM

| | | | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 9 | 8 | 9 | 9 | 5 | 2 | 6 | 2 | 2 | 4 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature *MB*

98995262240000001
West Nile Virus IgM
CONTROL 0.425 mL IVD
LOT 952
2024-07
DIESSE Monteriggioni Italy
CE
C4090

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Alciardi*

DATE *09/08/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.