

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME DENGUE IgG - CHORUS

Code	81154		
Lot No.	952	2024-07	2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D5730	Dengue IgG	952	2024-07	6X6
B5730	Calibratore Dengue IgG	952	2024-07	1
C5730	Control Dengue IgG	952	2024-07	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	650	≥ 100
Control (Index)	1,8	1,4-2,7

CALIBRATOR

1	0	8	9	5	2	6	0	0	0	0	0	0	4	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES  Operator Signature



CONTROL SERUM

1	0	8	9	5	2	6	1	4	2	7	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature



**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 21/08/2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.