



DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA PNEUMONIAE IgG - CHORUS

Code	81034		
Lot No.	952	2024-07	2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D2251	MYCOPLASMA PNEUMONIAE IgG	952	2024-07	6x6
B2252	CALIBRATORE MYCOPLASMA PNEUMONIAE IgG	952	2024-07	1
C2251	CONTROL SERUM MYCOPLASMA PNEUMONIAE IgG	952	2024-07	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	521	≥ 100
Control Serum (UA/mL)	28,8	17,5-38,0
Negative Control (UA/mL)	<10	<10

CALIBRATOR

0	6	6	9	5	2	6	8	1	4	2	4	5	7	5	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *HB*

066952681424575400
Mycoplasma pneumoniae IgG
CALIBRATOR 0.425 mL IVD
LOT 952
2024-07
2°C-8°C
DIESE Monteriggioni Italy CE
B2252

CONTROL SERUM

0	6	6	9	5	2	6	1	2	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *HB*

066952612230000001
Mycoplasma pneumoniae IgG
CONTROL+ 0.425 mL IVD
LOT 952
2024-07
2°C-8°C
DIESE Monteriggioni Italy CE
C2251

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *Bruno*

DATE *26/08/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive