

QUALITY CONTROL CERTIFICATE

KIT NAME tTg-G - CHORUS

| | | | |
|---------|-------|---------|---------|
| Code | 86060 | | |
| Lot No. | 952 | 2024-07 | 2023-07 |

REAGENTS

| Label | Description | Lot. | Exp. | NO |
|-------|---------------------|------|---------|-----|
| D3451 | tTg-G | 952 | 2024-07 | 6X6 |
| B3451 | CALIBRATORE tTg-G | 952 | 2024-07 | 1 |
| C3451 | CONTROL SERUM tTg-G | 952 | 2024-07 | 1 |


CONTROLS

| | Found results | Expected results |
|-----------------------|---------------|------------------|
| Calibrator (mOD) | 610 | ≥ 100 |
| Control Serum (AU/mL) | 23,5 | 15,0-49,1 |

CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 3 | 4 | 0 | 9 | 5 | 2 | 6 | 1 | 1 | 2 | 2 | 3 | 7 | 0 | 3 | 1 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature MB



CONTROL SERUM

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 3 | 4 | 0 | 9 | 5 | 2 | 6 | 1 | 0 | 1 | 9 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature MB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Beus

DATE 24/08/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive