

QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA PNEUMONIAE IgM - CHORUS

Code	81035		
Lot No.	945	2024-09	2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D2261	MYCOPLASMA PNEUMONIAE IgM	945	2024-09	6x6
B2261	CALIBRATORE MYCOPLASMA PNEUMONIAE IgM	945	2024-10	1
C2261	CONTROL SERUM MYCOPLASMA PNEUMONIAE IgM	945	2024-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	410	≥ 100
Control Serum (INDEX)	2,7	1,6-3,0
Negative Control (INDEX)	0,2	≤ 0,7

CALIBRATOR

0	6	7	9	4	5	7	0	0	0	0	0	0	4	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

06794670000430000
Mycoplasma pneumoniae IgM
CALIBRATOR 0.175 mL IVD
LOT 945
2024-10 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

0	6	7	9	4	5	7	1	6	3	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

067945716300000001
Mycoplasma pneumoniae IgM
CONTROL+ 0.425 mL IVD
LOT 945
2024-10 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 22/08/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.