



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME SM - CHORUS

Code	86016		
Lot No.	946	2024-09	2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3161	SM	946	2024-09	6X6
B3161	CALIBRATORE SM	946	2024-09	1
C3161	CONTROL SERUM SM	946	2024-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	572	≥ 100
Control Serum (AU/mL)	62,7	43,3-78,3

CALIBRATOR

1	9	2	9	4	6	7	3	1	5	2	4	3	7	4	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

192946731524374800
Sm
CALIBRATOR 0.175 mL IVD
LOT 946
 2024-09 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

1	9	2	9	4	6	7	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

192946719350000001
Sm
CONTROL+ 0.425 mL IVD
LOT 946
 2024-09 2°C-8°C
DIESSSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 30/08/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.