

QUALITY CONTROL CERTIFICATE

KIT NAME HELICOBACTER PYLORI IgG -CHORUS

Code	81060		
Lot No.	949	2024-06	2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D2391	Helicobacter Pylori IgG	949	2024-06	6X6
B2392	CALIBRATORE Helicobacter Pylori IgG	949	2024-06	1
C2391	CONTROL SERUM Helicobacter Pylori IgG	949	2024-06	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	410	≥ 100
Control Serum (UA/mL)	80,1	54,2-122,7
Negative Control (UA/mL)	<5	< 5

CALIBRATOR

0	0	8	9	4	9	6	2	1	9	3	1	3	6	5	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SB*



CONTROL SERUM

0	0	8	9	4	9	6	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 25.09.23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.