



QUALITY CONTROL CERTIFICATE

KIT NAME PSA

Code	86550		
Lot No.	970	2024-12	2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D2020	PSA	970	2024-12	6X6
B2020	CALIBRATOR PSA	970	2026-01	1
C2020	CONTROL SERUM PSA	970	2026-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	430	≥ 100
Control Serum (uIU/ml)	19,2	15,7-22,5

CALIBRATOR

9	4	8	9	7	0	7	1	4	4	5	2	2	9	6	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *RB*

948970714452296100
PSA
CALIBRATOR 0.250 mL IVD
LOT 970
2026-01 2°C-8°C
DIESE Monteriggioni Italy CE 0123

CONTROL SERUM

9	4	8	9	7	0	7	2	9	4	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *RB*

948970729400000001
PSA
CONTROL 0.500 mL IVD
LOT 970
2026-01 2°C-8°C
DIESE Monteriggioni Italy CE 0123

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *RB*

DATE 26/09/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.