



DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

QUALITY CONTROL CERTIFICATE

KIT NAME BRUCELLA IgG CHORUS

Code 81164
Lot No. 970 2024-09 2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D4110	Brucella IgG	970	2024-09	6X6
B4110	CALIBRATORE Brucella IgG	970	2024-09	1
C4110	CONTROL SERUM Brucella IgG	970	2024-09	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	550	≥ 100
Control (Index)	2,0	1,2-2,2

CALIBRATOR

1 4 6 9 7 0 6 0 0 0 0 0 5 8 0 0 0 0 0

Label checked with barcode reader YES Operator Signature

146970600000580000
Brucella IgG
CALIBRATOR 0,176 mL IVD
LOT 970
2024-09 2°C -8°C
DIESE Monteriggioni Italy CE

CONTROL SERUM

1 4 6 9 7 0 6 1 2 2 2 0 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature

146970612220000001
Brucella IgG
CONTROL+ 0,425 mL IVD
LOT 970
2024-09 2°C -8°C
DIESE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 26.08.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent,