



QUALITY CONTROL CERTIFICATE

KIT NAME MUMPS IgM CHORUS

Code	81076/12		
Lot No.	963	2024-08	2023-08

REAGENTS

Label	Description	Lot.	Exp.	NO
D2461	MUMPS IgM	963	2024-08	2x6
B2461	CALIBRATORE MUMPS IgM	963	2024-08	1
C2461	CONTROL SERUM MUMPS IgM	963	2024-08	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	420	≥ 100
Control Serum (Index)	2,3	1,8-3,3
Negative Control (Index)	0,0	< 0,7

CALIBRATOR

1	3	2	9	6	3	6	0	0	0	0	0	4	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *VB*

13296360000400000  
Mumps IgM  
CALIBRATOR 0.175 mL IVD  
LOT 963  
2024-08 2°C-8°C  
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

1	3	2	9	6	3	6	1	8	3	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *VB*

132963618330000001  
Mumps IgM  
CONTROL SERUM 0.425 mL IVD  
LOT 963  
2024-08 2°C-8°C  
DIESSSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Amie*

DATE 28/09/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.