



DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME RUBELLA IgG - CHORUS

| | | | |
|---------|-------|---------|---------|
| Code | 81030 | | |
| Lot No. | 954 | 2024-07 | 2023-07 |

REAGENTS

| Label | Description | Lot. | Exp. | NO |
|-------|---------------------------|------|---------|-----|
| D2231 | RUBELLA IgG | 954 | 2024-07 | 6X6 |
| B2232 | CALIBRATORE RUBELLA IgG | 954 | 2024-07 | 1 |
| C2231 | CONTROL SERUM RUBELLA IgG | 954 | 2024-07 | 1 |

CONTROLS

| | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

| | | |
|--------------------------|------|-----------|
| Calibrator (mOD) | 510 | ≥ 100 |
| Control Serum (IU/mL) | 25,0 | 18,0-37,5 |
| Negative Control (IU/mL) | <5 | < 5 |

081954662541404700
Rubella IgG
CALIBRATOR 0.425 mL IVD
LOT 954
2024-07
2°C-8°C
DIESE Monteriggioni Italy 0123

CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 8 | 1 | 9 | 5 | 4 | 6 | 6 | 2 | 5 | 4 | 1 | 4 | 0 | 4 | 7 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature *MB*

CONTROL SERUM

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 8 | 1 | 9 | 5 | 4 | 6 | 1 | 8 | 3 | 3 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature *MB*

081954618330000001
Rubella IgG
CONTROL+ 0.425 mL IVD
LOT 954
2024-07
2°C-8°C
DIESE Monteriggioni Italy 0123

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Agua*

DATE *27/09/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!