



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

QUALITY CONTROL CERTIFICATE

KIT NAME VARICELLA IgM CHORUS

Code	81079
Lot No.	954
	2024-10
	2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D2481	VARICELLA IgM	954	2024-10	6x6
B2481	CALIBRATORE VARICELLA IgM	954	2024-10	1
C2481	CONTROL SERUM VARICELLA IgM	954	2024-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	600	≥ 100
Control Serum (Index)	2,7	1,7-3,1
Negative Control (Index)	0,0	< 0,7

CALIBRATOR

1	3	4	9	5	4	7	0	0	0	0	0	4	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Osman*

13495470000420000
Varicella IgM
CALIBRATOR 0.175 mL IVD
LOT 954
2024-10 2°C-8°C
DIESSSE Monteriggioni Italy CE
B2481

CONTROL SERUM

1	3	4	9	5	4	7	1	7	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Osman*

13495471731000001
Varicella IgM
CONTROL+ 0.425 mL IVD
LOT 954
2024-10 2°C-8°C
DIESSSE Monteriggioni Italy CE
C2481

This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Osman*

DATE 27-09-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.