

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** ENA-6 S - CHORUS

Code	86012		
Lot No.	963	2024-08	2023-08

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D3131	ENA-6 S	963	2024-08	6X6
B3131	CALIBRATORE ENA-6 S	963	2024-08	1
C3131	CONTROL SERUM ENA-6 S	963	2024-08	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	406	≥ 100
Control Serum (ndx)	2,7	1,9-3,5

**CALIBRATOR**

1	0	2	9	6	3	6	0	0	0	0	0	3	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

10296360000370000  
ENA-6 S  
CALIBRATOR 0.175 mL IVD  
LOT 963  
 2024-08 2°C-8°C  
DIESSE Monteriggioni Italy CE  
B3131

**CONTROL SERUM**

1	0	2	9	6	3	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

102963619350000001  
ENA-6 S  
CONTROL 0.425 mL IVD  
LOT 963  
 2024-08 2°C-8°C  
DIESSE Monteriggioni Italy CE  
C3131

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 03.10.2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.