



QUALITY CONTROL CERTIFICATE

KIT NAME SS-A - CHORUS

Code	86018			
Lot No.	970	2024-12	2023-09	

REAGENTS

Label	Description	Lot.	Exp.	NO
D3181	SS-A	970	2024-12	6X6
B3181	CALIBRATORE SS-A	970	2024-12	1
C3181	CONTROL SERUM SS-A	970	2024-12	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	390	≥ 100
Control Serum (AU/mL)	38,2	23,3-57,0

CALIBRATOR

1	9	3	9	7	0	7	1	1	9	2	9	4	7	4	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature HB



CONTROL SERUM

1	9	3	9	7	0	7	1	5	2	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature HB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. HB

DATE 06/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive