



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME ANA-Screen - CHORUS

Code	86014		
Lot No.	971	2024-09	2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D3141	ANA-Screen	971	2024-09	6X6
B3141	CALIBRATORE ANA-Screen	971	2024-09	1
C3141	CONTROL SERUM ANA-Screen	971	2024-09	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	248	≥ 100
Control Serum (Ndx)	3,7	2,6-4,9

CALIBRATOR

1	1	5	9	7	1	6	0	0	0	0	0	2	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Bohler*



CONTROL SERUM

1	1	5	9	7	1	6	2	6	4	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Bohler*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *Bohler*

DATE 13.10.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.