

QUALITY CONTROL CERTIFICATE

KIT NAME Chorus Chlamydia trachomatis IgA

Code	81262
Lot No.	952
	2024-07
	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D9650	Chorus Chlamydia trachomatis IgA	952	2024-07	6X6
B9650	CALIBRATOR Chlamydia trachomatis IgA	952	2024-09	1
C9650	POSITIVE CONTROL Chlamydia trachomatis IgA	952	2024-09	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	479	≥ 100
Positive Control (Index)	1,9	1,6-2,9

CALIBRATOR

2	3	2	9	5	2	6	0	0	0	0	0	4	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Alber*

23295260000450000
Chlamydia trachomatis IgA
CALIBRATOR 0.175 mL IVD
LOT 952
2024-09 2°C 8°C
DIESE Monteriggioni Italy CE 0123 B9650

CONTROL SERUM

2	3	2	9	5	2	6	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Alber*

232952616290000001
Chlamydia trachomatis IgA
CONTROL 0.425 mL IVD
LOT 952
2024-09 2°C 8°C
DIESE Monteriggioni Italy CE 0123 C9650

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Alber*

DATE 13-10-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.