



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME LEGIONELLA PNEUMOPHILA 1 IgG -CHORUS

Code	81091		
Lot No.	978	2024-07	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D7371	LEGIONELLA PNEUMOPHILA 1 IgG	978	2024-07	6X6
B7371	CALIBRATORE L. PNEUMOPHILA 1 IgG	978	2024-07	1
C7371	CONTROL SERUM L. PNEUMOPHILA 1 IgG	978	2024-07	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	430	≥ 100
Control Serum (INDEX)	1,9	1,4-2,6
Negative Control (INDEX)	0,0	<0,7

CALIBRATOR

1	5	5	9	7	8	5	0	0	0	0	0	5	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature IB



CONTROL SERUM

1	5	5	9	7	8	5	1	4	2	6	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature IB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents IB

DATE 25-10-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.