

QUALITY CONTROL CERTIFICATE

KIT NAME DENGUE IgG - CHORUS

|         |         |
|---------|---------|
| Code    | 81154   |
| Lot No. | 979     |
|         | 2024-10 |
|         | 2023-10 |

REAGENTS

| Label | Description            | Lot. | Exp.    | NO  |
|-------|------------------------|------|---------|-----|
| D5730 | Dengue IgG             | 979  | 2024-10 | 6X6 |
| B5730 | Calibratore Dengue IgG | 979  | 2024-10 | 1   |
| C5730 | Control Dengue IgG     | 979  | 2024-10 | 1   |

CONTROLS

|                  | Found results | Expected results |
|------------------|---------------|------------------|
| Calibrator (mOD) | 586           | ≥ 100            |
| Control (Index)  | 2,3           | 1,7-3,2          |

CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 0 | 8 | 9 | 7 | 9 | 6 | 0 | 0 | 0 | 0 | 0 | 0 | 4 | 1 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES  Operator Signature *[Signature]*



CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 1 | 0 | 8 | 9 | 7 | 9 | 6 | 1 | 7 | 3 | 2 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES  Operator Signature *[Signature]*



*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 25-10-2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.