



QUALITY CONTROL CERTIFICATE

KIT NAME tTg-A - CHORUS

Code	86058		
Lot No.	959	2024-07	2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D3431	tTg-A	959	2024-07	6X6
B3431	CALIBRATORE tTg-A	959	2024-07	1
C3431	CONTROL SERUM tTg-A	959	2024-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	458	≥ 100
Control Serum (AU/mL)	21,1	19,3-36,4

CALIBRATOR

3	3	9	9	5	9	6	3	2	4	3	8	5	0	4	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

339959632436504500
tTg-A
CALIBRATOR 0.178 mL IVD
LOT 959
2024-07
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

3	3	9	9	5	9	6	1	4	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

339959614270000001
tTg-A
CONTROL+ 0.425 mL IVD
LOT 959
2024-07
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 27/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.