



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME CYTOMEGALOVIRUS IgG -CHORUS

Code	81010		
Lot No.	978	2024-10	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D2161	Cytomegalovirus IgG	978	2024-10	6x6
B2162	CALIBRATORE Cytomegalovirus IgG	978	2024-10	1
C2161	CONTROL SERUM Cytomegalovirus IgG	978	2024-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	390	≥ 100
Control Serum (IU/mL)	9,5	5,0-15,0
Negative Control (IU/mL)	<0,5	<0,5

CALIBRATOR

0	0	1	9	7	8	6	6	2	5	3	6	2	8	5	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

001978662536285600
Cytomegalovirus IgG
CALIBRATOR 0.425 mL IVD
LOT 978
2024-10 2°C-8°C
DIESSE Monteriggioni Italy CE 0123 B2162

CONTROL SERUM

0	0	1	9	7	8	6	2	5	4	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

001978625460000001
Cytomegalovirus IgG
CONTROL SERUM 0.425 mL IVD
LOT 978
2024-10 2°C-8°C
DIESSE Monteriggioni Italy CE 0123 B2162

This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents MB

DATE 31.10.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.