



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME 88000 TOTAL IgE

Code	88000		
Lot No.	980	2024-10	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D5771	88000- TOTAL IgE	980	2024-10	6x6
B5771	CALIBRATORE 88000- TOTAL IgE	980	2025-01	1
C5771	CONTROL SERUM 88000- TOTAL IgE	980	2025-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	280	≥ 100
Control Serum (KIU/l)	218,2	140,9-343,7

CALIBRATOR

9	1	1	9	8	0	6	3	4	3	5	1	2	8	6	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

911980634351286500
Total IgE
CALIBRATOR 0.400 mL IVD
LOT 980
2025-01 2°C-8°C B5771
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

9	1	1	9	8	0	6	2	5	4	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

911980625460000001
Total IgE
CONTROL+ 0.850 mL IVD
LOT 980
2025-01 2°C-8°C C5771
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 31.10.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.