

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F83 CHICKEN MEAT

Code	88000-F83		
Lot No.	985	2024-10	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D5701	88000-F83 CHICKEN MEAT	985	2024-10	2X6
B5700	CALIBRATORE 88000-F83 CHICKEN MEAT	985	2024-10	1
C5700	CONTROL SERUM 88000-F83 CHICKEN MEAT	985	2024-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	267	≥ 100
Control Serum (KU/L)	9,5	7,5-15,7

CALIBRATOR

6	5	0	9	8	5	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



CONTROL SERUM

6	5	0	9	8	5	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 02.11.23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.