



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME GLIADIN-A - CHORUS

Code	86054		
Lot No.	978	2024-10	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D3391	Gliadin-A	978	2024-10	6X6
B3392	CALIBRATORE Gliadin-A	978	2024-10	1
C3392	CONTROL SERUM Gliadin-A	978	2024-10	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	481	≥ 100
CONTROL SERUM (AU/ml)	71,2	50,0-95,2

CALIBRATOR

5	0	1	9	7	8	6	3	2	3	4	4	3	1	6	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature MB



CONTROL SERUM

5	0	1	9	7	8	6	2	3	4	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Agnese

DATE 03/11/23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.