



QUALITY CONTROL CERTIFICATE

KIT NAME PSA

Code	86550		
Lot No.	978	2025-01	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D2020	PSA	978	2025-01	6X6
B2020	CALIBRATOR PSA	978	2026-01	1
C2020	CONTROL SERUM PSA	978	2026-01	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	4,30	≥ 100
Control Serum (uIU/ml)	16,7	14,8-20,2

CALIBRATOR

9	4	8	9	7	8	7	1	4	5	5	3	3	1	6	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB



CONTROL SERUM

9	4	8	9	7	8	7	2	8	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB



*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents... FB

DATE 02/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.