



QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS DENGUE IgM CAPTURE

Code	81157		
Lot No.	970	2024-09	2023-09

REAGENTS

Label	Description	Lot.	Exp	No
D4140	Dengue IgM Capture	970	2024-09	6X6
B4140	Calibratore Dengue IgM Capture	970	2024-09	1
C4140	Control Dengue IgM Capture	970	2024-09	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	400	≥ 100
Control (Index)	3,3	2,4-4,4

CALIBRATOR

9	8	6	9	7	0	6	0	0	0	0	0	3	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

98697060000300000
Dengue IgM Capture
CALIBRATOR 0.450 mL IVD
LOT 970
2024-09 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

9	8	6	9	7	0	6	2	4	4	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

98697062444000001
Dengue IgM Capture
CONTROL+ 0.900 mL IVD
LOT 970
2024-09 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 02.11.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.