



QUALITY CONTROL CERTIFICATE

KIT NAME SM - CHORUS

Code	86016		
Lot No.	948	2024-09	2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3161	SM	948	2024-09	6X6
B3161	CALIBRATORE SM	948	2024-09	1
C3161	CONTROL SERUM SM	948	2024-09	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	550	≥ 100
Control Serum (AU/mL)	68,2	51,8-82,9

CALIBRATOR

1	9	2	9	4	8	7	3	1	4	2	5	3	2	5	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SB*

192948731425325300
Sm
CALIBRATOR 0.175 mL [IVD]
LOT 948
2024-09 2°C B3161
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

1	9	2	9	4	8	7	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SB*

192948722410000001
Sm
CONTROL+ 0.425 mL [IVD]
LOT 948
2024-09 2°C C3161
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*
DATE *09.11.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.