



QUALITY CONTROL CERTIFICATE

KIT NAME RUBELLA IgG - CHORUS

Code	81030		
Lot No.	979	2024-10	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D2231	RUBELLA IgG	979	2024-10	6X6
B2232	CALIBRATORE RUBELLA IgG	979	2024-10	1
C2231	CONTROL SERUM RUBELLA IgG	979	2024-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	577	≥ 100
Control Serum (IU/mL)	20,0	14,0-27,5
Negative Control (IU/mL)	<5	< 5

CALIBRATOR

0	8	1	9	7	9	6	7	2	5	3	5	5	1	4	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

081979672535514500
Rubella IgG
CALIBRATOR 0.425 mL IVD
LOT 979
2024-10 2°C-8°C
DIESSE Monteriggioni Italy
B2232
CE 0123

CONTROL SERUM

0	8	1	9	7	9	6	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

081979614260000001
Rubella IgG
CONTROL+ 0.425 mL IVD
LOT 979
2024-10 2°C-8°C
DIESSE Monteriggioni Italy
C2231
CE 0123

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Leone

DATE 09/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.