



QUALITY CONTROL CERTIFICATE

KIT NAME Vitamin D Total- CHORUS

Code	86900		
Lot No.	954	2024-07	2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D9702	25OH Vitamin D Total	954	2024-07	6x6
B9703	CALIBRATOR 25OH Vitamin D Total	954	2025-05	1
C9703	CONTROL SERUM 25OH Vitamin D Total	954	2025-05	4


CONTROLS

	Found results	Expected results
Calibrator (mOD)	471	≥ 100
CONTROL SERUM (ng/ml)	38,6	32,4-39,3

CALIBRATOR

8	3	5	9	5	4	6	1	3	6	4	0	6	3	6	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *SB*



CONTROL SERUM

8	3	5	9	5	4	6	1	1	2	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *SB*

DATE *09.11.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.