

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-M - CHORUS

Code	86034		
Lot No.	978	2025-01	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D7401	dsDNA-M	978	2025-01	6X6
B7401	CALIBRATORE dsDNA-M	978	2025-01	1
C7401	CONTROL SERUM dsDNA-M	978	2025-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	560	≥ 100
Control Serum (AU/mL)	59,7	39,6-87,5

CALIBRATOR

2	5	7	9	7	8	7	5	2	2	3	4	4	2	4	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

257978752234424600
dsDNA-M
CALIBRATOR 0,176 mL IVD
LOT 978
 2025-01 2°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

2	5	7	9	7	8	7	1	7	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

257978717310000001
dsDNA-M
CONTROL 0,426 mL IVD
LOT 978
 2025-01 2°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 07.11.23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.