



QUALITY CONTROL CERTIFICATE

KIT NAME BORRELIA IgG - CHORUS

Code	81046		
Lot No.	988	2024-11	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D2300	BORRELIA IgG	988	2024-11	6X6
B2300	CALIBRATORE BORRELIA IgG	988	2025-02	1
C2300	CONTROL BORRELIA IgG	988	2025-02	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	380	≥ 100
Control (Index)	2,9	2,0-3,7

CALIBRATOR

9	0	4	9	8	8	6	0	0	0	0	0	3	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*



CONTROL SERUM

9	0	4	9	8	8	6	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Benoc*

DATE *08.11.23*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.