



DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME ft3 - CHORUS

Code	86518		
Lot No.	987	2024-11	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D3300	ft3	987	2024-11	6X6
B3300	CALIBRATOR ft3	987	2024-11	1
C3300	CONTROL SERUM ft3	987	2024-11	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	331	≥ 100
Control Serum (pmol/L)	3,9	2,1-8,0

CALIBRATOR

9	6	7	9	8	7	6	5	3	0	4	0	2	7	6	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature FB



CONTROL SERUM

9	6	7	9	8	7	6	2	6	4	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Manelf

DATE 15/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.